

**वी.पी.कोइराला स्वास्थ्य विज्ञान प्रतिष्ठान, धरानको
कोटेशन माग गरिएको सूचना**

(पुनः सूचना प्रकाशित मिति २०८०/०६/१७)

यस प्रतिष्ठानको अस्पताल फार्मेसीमा बिक्री बितरण प्रयोजनको लागि (Drug Eluting Stents) आवश्यक भएकोले बोलपत्र Tender को माध्यमबाट खरिद गर्नुपर्ने भएकोले इजाजत प्राप्त प्रतिष्ठानमा सुचिकृत फर्म, संस्था, कम्पनीबाट अद्यावधिक फर्म दर्ता प्रमाणपत्र, मु.अ.कर दर्ता प्रमाणपत्र, आ.व.२०७९।०८० को कर चुक्ताको प्रमाणपत्र प्रतिहरू संलग्न गरी यो सूचना प्रकाशित मितिले ७ दिन भित्र कोटेशन पेश गर्नुहुन सुचित गरिन्छ । माग गरिएको कोटेशन तोकिएको म्यादभित्र प्रतिष्ठानको दर्ता चलानी फांटमा दर्ता गर्नुहुन वा quotation.procurement@bpkihs.edu मा Email मार्फत पेश गर्नुपर्नेछ ।

Drug Eluting Stents

S.N.	Description	Unit	QTY
1	(USFDA approved latest generation Everolimus+cobalt chromium stent (xience Xpedition or later generation DES) Stent of all Available Diameters and Different Lengths. All Product Must Have Expiry date limit of at least 1 year from date of Supply)	Set	30
2	(USFDA approved Latest Generation Zotarolimus + Cobalt chromium stent (Resolute integrity or later generation DES). Stents Of all Available Diameters and Different Lengths. All product must have Expiry date limit of at least 1 year from date of supply.)	Set	30
3	(USFDA approved Latest Generation sirolumis + Cobalt With Biodegradable Polymer and Ultra thin 60-80µm Stent struts. hromium stent) . Stents Of all Available Diameters and Different Lengths. All product must have Expiry date limit of at least 1 year from date of supply with Probio Passive Coating and Biodegradeble Polymer.)	Set	30
4	(USFDA approved Latest Generation Everolimus + Chromium Platinum stents (PROMUS ELEMENT PLUS or Later Generation DES). Stents Of all Available Diameters and Different Lengths. All product must have Expiry date limit of at least 1 year from date of supply.)	Set	30